

A

B

C

D

Optical Specification

Model	OPT-CH08-65C-F22
Optical Configuration	Telecentric
Working Distance (mm)	65±2
Magnification	0.8X
Sensor Ø(mm)	11.0(2/3")
Aperture	F22
Telecentricity	0.10°
NA	0.01818
Distortion	<0.1%
Depth of Field (mm)	2.84
Object Resolution (μm)	18.45
Resolution (μm)	11.10
Image MTF30(lp/mm)	45
Field of View Ø(mm)	13.75
O/I Distance (mm)	191.6
Mount	C-Mount
Weight (kg)	0.14
Size (mm)	Ø29.0×109.06
Filter Thread (mm)	M27×0.5-6H
Working temperature	-20°C~+60°C

[1]Resolution:The resolution is the image resolution.

[2]WD:The distance between the front end of the mechanism and the object.

[3]DOF evaluation formula for telecentric lenses:

$$\text{DOF(mm)} = \frac{F\# \times P \times k}{\beta \times \beta}$$

F#:The real F-number of a lens when used


P:Sensor pixel size(5.5 μm used here)

β:Lens magnification


k:The parameter based on application (measurement,

k=0.015;defect detection, k=0.008;0.015 used here)

At the borders of DOF,the image can still be used for measurement. But only half of the nominal DOF should be used to get a sharp image.



View Angle



Item

Signature

Date



OPT Machine Vision Tech Co., Ltd.

Standard Difference

IT14

Drawing

Design

Check

All Rights Reserved

Drawing Code

Scale

Paper Size

A4

Name

OPT-CH08-65C-F22

Unit

mm

Page 1 / 1

Version

1.0

A

B

C

D